

Warning!

NEVER USE ACETONE! The use of certain solvents on 4-hole distal attachment blocks could reduce their strength resulting in structural failure of the prosthesis. If you are going to use a solvent to clean the blocks only use alcohol and **never use acetone.**

This warning refers to distal attachment block kits 124102 and FC100150.

PM167/04-03